

Vol.05 No.2

Prevention of accidental retention of vaginal swabs and tampons during childbirth using a novel 'design' solution within a 'human factors systems approach' towards patient safety

Aditi Desai University of Keele, UK



Abstract

Introduction: In UK, retained swabs after vaginal birth and perineal suturing are classed as "Never events.1 Retained vaginal swabs have accounted for around 30% of retained foreign object -'never events' reported on an annual basis for a number of years. Retained vaginal swab post-delivery could cause significant morbidity and patient may experience serious physical and psychological complications, including infection, secondary post-partum haemorrhage, sepsis, depression, lack of bonding with their baby and loss of trust in the healthcare system. As yet, the interventions developed to prevent this problem have focussed towards changing human behaviour. As per hazard control hierarchy, a 'human factors' engineered solution is more likely to be effective than just modifying human behaviour4. We present a 'human factors engineered system' where a design solution and human behaviours complement each other, to prevent human error.

Aim of the project: To prevent the accidental retention of vaginal swab with a simple device so as to 'design out' the human error. We present findings from a formative usability study done in a high fidelity simulation environment, which looked at the user experience, safety and acceptability of the novel device system.

Description of the device: The device has a seatbelt plug in system where by each swab tail has a plastic clip at the end which plugs into the device displaying a green indicator when all are plugged in and a red indicator when one or more are disconnected. The aim is to ensure that all swabs are returned and not retained in the patient. Its major benefit is its simplicity and cost effectiveness.

Methods: Seven healthcare professionals who work in that particular clinical area participated in the usability study done by NIHR Medtech, University hospitals, Birmingham. Each user session lasted for approximately 40 minutes and it involved audio video content, hands-on session using the device in a high fidelity theatre simulation set up and feedback in the form of a semistructured interview

Usability study findings: All the participants have a positive attitude towards the device being supportive of the concept but made some valuable design recommendations based on their clinical practice and personal preferences. The device is external to the patient and is mechanical and therefore poses no risk to the user or the patient as observed in the simulated environment. The feedback suggested improving the design of the clips in order to make them finer and to make the device compact and easy to handle. This valuable feedback led to further prototype iterations in the device development journey. A simulation study and a pilot study in childbirth clinical environment is being planned to develop further understanding.

Conclusion: The users found the design solution acceptable and felt positive about the concept. Further evaluation studies are required to establish the reliability of the system in preventing the accidental retention. Further evaluation studies should be planned to establish the reliability of the system in preventing accidental retention and to design a 'human factors engineered system' where a design solution and human behaviours complement each other, to prevent human error.



Biography:

Vol.05 No.2



Experienced Consultant Obstetrician and Gynaecologist with a demonstrated history of working in the hospital & health care industry since 1995. Skilled in Obstetrics and Gynaecology, Change Management in healthcare, Clinical research and Social enterprise. Holds the fellowship of the Royal College of Obstetricians and Gynaecologists for contribution to the field (2015) .Also holds a leadership and management qualification from the NHS Leadership academy. Member of the Chartered Institute of Ergonomics and Human factors Interested in Innovation, Human Factors and Patient safety. Senior clinical examiner for year 5, MBChB examinations, Birmingham University

Speaker Publications:

1. NHS England. Never Events List 2018/2019.

2. NHS Improvement. Never Events reported as occurring between 1 April 2017 and 31 March 2018 – final update. 2019.

3. Mahran, M. A., Toeima, E. & Morris, E. P. (2013) The recurring problem of retained swabs and instrument. Best Practice and Research: Clinical Obstetrics and Gynaecology, 27 (4): 489-495.

4. Health and Safety Executive. (2019) frequently asked questions- what is the hierarchy of control? [Online] Available at: http://www.hse.gov.uk/risk/faq.htm#hierarchy

<u>5th World Congress on Patient Safety & Quality Healthcare</u>, June 22-23, 2020.

Abstract Citation:

Aditi Desai, Prevention of accidental retention of vaginal swabs and tampons during childbirth using a novel 'design' solution within a 'human factors systems approach' towards patient safety, June 22-23, 2020 (https://patientsafety.insightconferences.com/speaker/2020/aditi -desai-university-of-keele-uk)