

## Suture Button Fixation for Syndesmosis Injuries: Review of the Literature

Meric Unal A\*

Orthopaedics and Traumatology, Isparta Sifa Hospital, Isparta, Turkey

\*Corresponding author: Meric Unal A, Orthopaedics and Traumatology, Isparta Sifa Hospital, Isparta, Turkey, Tel: 0090 507 0240904; E-mail: [abdmunal@yahoo.com](mailto:abdmunal@yahoo.com)

Received date: Jan 28, 2014, Accepted date: May 28, 2014, Published date: Jun 07, 2014

Copyright: © 2014 Unal MA. This is an open-access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

### Abstract

Syndesmosis injuries -also known as high ankle sprains- have an increasing popularity because of an increased awareness of diagnosis. Different from lateral ankle sprains, mechanism of injury mostly involves external rotation, eversion and excessive dorsiflexion. Other mechanisms can also cause syndesmosis injuries. Because of high levels of missed injuries the real prevalence of syndesmosis injuries is underestimated. It is less frequently seen than lateral ankle injuries. For diagnosis, physical and radiological examination is essential.

In radiological examination, associated injuries must be determined if present. Mostly seen associated injury is ankle fracture. Nonoperative treatment provides good results for syndesmosis injuries. Surgical treatment is indicated when syndesmosis injuries include frank diastasis. Arthroscopic view of syndesmotic instability is another indication. Failed conservative treatment may also be an indication. There are a lot of options for surgical fixation. Screws are the most popular among others but there are some complications about them. Alternatively, suture button fixation technique can be used. Suture button fixation is an implant with two metallic buttons surrounded with thick fiberwire sutures. Its stabilization mechanism depends on compression of two buttons to opposite sites by the help of fiberwire. It is designed to resist diastasis but allows small movement to other planes. It is very suitable for fixation of syndesmosis injuries and also to have less complication rates. Purpose of this study is to discuss the syndesmotic injuries and review the suture button fixation technique for the treatment.

**Keywords:** Syndesmosis injuries; Ankle sprains; Antero inferior tibiofibular ligament; Osteochondral

### Introduction

Syndesmosis injuries, also known as high ankle sprains, have an increasing popularity because of an increased awareness of diagnosis [1]. Injury spectrum includes syndesmotic ligament and interosseous membrane injuries and ankle fractures with syndesmotic disruption [2]. Because of high levels of missed injuries the real prevalence of syndesmosis injuries is underestimated [1]. It is less frequently seen than lateral ankle injuries [3]. The incidence differs between 1% to 18% in different studies [4-9].

### Anatomy and Mechanics

Syndesmosis composed of tibia and fibula. It is a very stable joint however fibular movement occurs to accommodate talus during gait [10]. Tibia and fibula are connected to each other with four major ligamentous structures. Anteriorly, the antero inferior tibiofibular ligament (AITFL) lies and it provides 35% of overall ankle stability. Second one is the interosseous ligament (IOL). It is the shortest one but the primary attachment between tibia and fibula. It also provides 22% of stability. The third one is the interosseous membrane (IOM) found at the superior aspect of IOL. It prevents posterolateral bowing of fibula and play role in load sharing ability of fibula. The fourth one lies posteriorly and named as posterior inferior tibiofibular ligament (PITFL). It has two portions known as superficial and deep (transverse tibiofibular ligament). Deep portion provides 33% and superficial portion provides 9% of ankle stability. Failure of two of these ligaments leads to mechanical laxity of syndesmosis [2,3,8,10,11].

Syndesmosis widens 1.5 mm physiologically while ankle dorsiflexion [8]. It is showed that failure of total PITFL leads the most syndesmosis widening among other ligaments [10,12].

### Mechanisms of Injury

Injuries to these four ligamentous structures lead to separation of syndesmosis. Different from lateral ankle sprains, mechanism of injury mostly involves external rotation, eversion and excessive dorsiflexion [1,3]. Less commonly seen mechanisms are dorsiflexion with axial loading, eversion, inversion, plantar flexion, pronation and internal rotation [4,7,13-17]. In most complete syndesmotic disruptions external rotation causes a Weber B or C fracture with syndesmosis widening [18,19]. When the ankle is in neutral position, excessive external rotation causes injury to only tibiofibular ligaments without damaging other structures. External rotation mechanism mostly seen in American football and skiing [20,21]. Hyperdorsiflexion mechanism is mostly seen in running and jumping like sports [4].

### Diagnosis

Diagnosis of syndesmotic injuries are based on careful physical examination [2,22]. Pain and tenderness on the anterior aspect of syndesmosis are the most common findings [22]. In pure ligamentous injuries, squeeze test (Hopkins test) and external rotation under stress test (Kleiger test) may be useful [1,3,9,10,22]. External rotation stress test found more reliable than squeeze test [1]. There are some other tests uncommonly performed like cotton test, cross leg test and stabilization test [9,10].

Radiological examination always starts with antero-posterior (AP), lateral and mortise views of ankle [1,2,22]. Direct radiology usually

used to rule out ankle fractures and detect syndesmotom separation [10]. The criteria for syndesmotom distruption are increased tibiofibular clear space, decreased tibiofibular overlap and increased medial clear space. Tibiofibular clear space should be < 6 mm in both AP and mortise views. Tibiofibular overlap should be > 6 mm in AP and > 1 cm in mortise view. Medial clear space is equal to or less than superior clear space between talar dome and tibial plafond [9,10,22] (Figure 1A and 1B). All injuries out of normal values for these measurements indicate syndesmotom injury. Stress radiographs are also useful for diagnosis [22].

Computed tomography (CT), magnetic resonance imaging (MRI) and arthroscopy are the advanced diagnostic methods for syndesmotom injuries [1,9,10,22]. CT can detect minor separations. MRI is highly sensitive and specific [22]. MRI also shows associated injuries like osteochondral lesions [1]. Bone scintigraphy is also can be used for detection of occult syndesmotom injuries. Arthroscopy allows clear vision to injury site and provides definite diagnosis for syndesmotom disruption [3,8]. Arthroscopy is a good support for accurate diagnosis.




Figure 1: Post-operative antero-postoral; 1B: post-operative lateral.

## Treatment

Syndesmotom injuries without fracture and separation can be treated conservatively [3,8,23]. Initial treatment starts with rest, ice, elevation and nonsteroidal anti-inflammatory drugs [3,23]. Immobilization like bracing or casting can be added to treatment according to degree of pain. Recovery time is prolonged in conservative treatment [8]. After one to six weeks rehabilitation can be started to restore function and strength of ankle [9,23]. Non-operative treatment provides good results over 86% of all syndesmotom injuries [23,24].

Surgical treatment is indicated when syndesmotom injuries include frank diastasis. Arthroscopic view of syndesmotom instability is another indication [10]. Failed conservative treatment may also be an indication.

Syndesmotom reduction and transsyndesmotom fixation is performed for standart surgical procedure [10]. Metallic screws are used for fixation implant classically. Besides metallic screws there are some alternatives for fixation which are; bioabsorbable screws, syndesmotom staple, ilizarov ring fixator, kirchner wires, flexible implants (suture buttons), syndesmotom hook, trans-syndesmotom bolt, cerclage wires, ANK nail or ligamentoplasties [25-39].

There is no agreement on number of screws, size of screws, optimal number of cortices, position of screws relative to tibiotalar joint, weight bearing time after surgery [24]. There are some problems reported about conventional screw fixation. These are; screw loosening, screw breakage, stiffness, prolonged non weight bearing for protection of screw, synositis, second operation for screw removal, late diastasis after screw removal because of insufficient fixation time [40-44].

Because of all these problems, some alternative fixation implants used or newly developed as listed above. The aims of the alternative fixation methods are; stable and flexible fixation, no need for second surgery, early weight bearing and functional recovery and less post-operative complications.

## Suture Button Fixation

Suture button fixation is an implant with two metallic buttons surrounded with thick fiber wire sutures. Its stabilization mechanism depends on compression of two buttons to opposite sites by the help of fiber wire. It is designed to resist diastasis but allows small movement to other planes. It can be performed as an open technique in syndesmotom with ankle fractures, also can be performed percutaneously in isolated syndesmotom seperations. The studies about suture button started with cadaver study on 2003 and after this, clinical trials, fixation comparisons and reviews were published.

Thornes et al. in a cadaver study compared external rotation strength of four cortices screw and suture button fixations. There is no significant difference between groups in loading test. Suture button give more consistent performance than screw [31].

Thornes et al. in their study; compared suture button fixation and screw fixation techniques in 16 patients retrospectively. Suture button fixation had better AOFAS scores and they return to work earlier than screw group. It is an easy technique, safe, effective and has high patient satisfaction. It is also found cost effective because there is no need for second surgery for implant removal [32].

Cottom et al. reported 25 cases of syndesmotom separation treated with suture button fixation (Tight-Rope, Arthrex Inc.) technique. They found this technique as a valid, safe and cost effective. They also recommended; performing of two suture buttons may be more useful for in Weber C or communicated Weber B type fibula fractures [45].

Thornes and McCartan published a case review series treated with suture button (Tight-Rope, Arthrex Inc.) at 2006. They found no loss of reduction, no major complications and no patient underwent to second surgery for implant removal. They recommended that joint dislocation, age > 50 and female sex are all associated with poor outcome [46].

Coetzee and Ebeling, in their study; evaluate the indications of suture button. In their prospective serie suture button (Tight-Rope, Arthrex Inc.) group had similar results when compared to screw group

but suture button group had better range of motion measurement, less stiffness and less discomfort [47].

Soin et al. in their biomechanical study; compare loading differences of suture button and screw fixation. They found similar results between groups in cyclic loading. Neither of the groups could not restore the native ankle motion [48].

In another biomechanical study, Klitzman et al. determine more physiologic fibular movement in sagittal plane at suture button fixation than screw fixation. They concluded that flexible fixation like suture button provides more physiologic healing to syndesmosis [49].

Degroot et al. published outcomes of suture button fixation. They recommend that suture button fixation needs more second surgery for implant removal than known before. Osteolysis may occur near the button and subsidence of implant may occur [50].

Teramoto et al. compared different suture button fixation techniques that are single, double and anatomical suture button fixations in cadaver specimens. Finally they said, neither of single or double suture button fixation could not provide multidirectional stability but anatomical fixation directed from posterior cortex of fibula to anterolateral edge of tibia provides dynamic stabilization [51].

Rigby and Cottom published their two year follow-up with suture button fixation (Tight-Rope, Arthrex Inc.). They concluded that this technique provides long term stability of ankle mortise [52].

In two review studies in 2012; Suture button fixation technique found similar to screw fixation but it has more advantages like it is less invasive, easily performed, safe and effective for syndesmosis separations. There is also less need to second surgery for implant removal and earlier return to work [21,53].

Storey et al. determined the complications of suture button syndesmosis. Fiber wire skin irritation, osteolytic reaction, rediastasis are the most seen complications. These complications could be prevented by Laying the fiberwire flat, remove the implant in case of osteolysis and make a mini medial incision for improving the stability of medial button for rediastasis [54].

The author's technique includes percutaneous suture button fixation for isolated syndesmosis injuries. If associated with Weber C fibula fractures the preferred fixation with the suture button is to perform it distal to the plate fixation of fracture mostly (Figure 2A and 2B). If associated with Weber B fibula fracture the suture button usually performed from the screw hole of the plate at the level of 1 cm proximal to syndesmosis [55,56].

## Conclusion

Among all these fixation techniques for syndesmosis injuries, suture button fixation has a lot of advantages besides its complications. This technique is less invasive, it may be performed percutaneously for isolated injuries. It needs less second surgery-for implant removal-among all other techniques so it has less morbidity to patient and it is more cost effective. It can be easily performed, technique is easy and practical because it has a sterile application kit. It is a very physiological fixation method because it provides flexible fixation and this also allows to patient early return to work. Besides all these advantages this technique also has high patient satisfaction.

Although suture button fixation has some complications, they are seen less than complications about other techniques and usually about

metallic and fiber wire based complications. They are all preventable and correctable. Fiber wire skin irritation, osteolytic reaction and rediastasis are mostly seen complications.

Among all these studies suture button fixation seems a very good choice for syndesmosis injuries. It may be the first fixation choice for many orthopaedic surgeons.



Figure 2: Pre-operative antero- postoral 2B: Post-operative lateral

## References

1. Peña FA, Coetzee JC (2006) Ankle syndesmosis injuries. *Foot Ankle Clin* 11: 35-50, viii.
2. Wuest TK (1997) Injuries to the Distal Lower Extremity Syndesmosis. *J Am Acad Orthop Surg* 5: 172-181.
3. Lin CF, Gross MT, Weinhold P (2006) Ankle Syndesmosis Injuries: Anatomy, Biomechanism of Injury, and Clinical Guidelines for Diagnosis and Intervention. *J Orthop Sports Phys Ther*: 372-384
4. Norkus SA, Floyd RT (2001) The anatomy and mechanisms of syndesmotic ankle sprains. *J Athl Train* 36: 68-73.
5. Fallat L, Grimm DJ, Saracco JA (1998) Sprained ankle syndrome: prevalence and analysis of 639 acute injuries. *J Foot Ankle Surg* 37: 280-285.
6. Gerber JP, Williams GN, Scoville CR, Arciero RA, Taylor DC (1998) Persistent disability associated with ankle sprains: a prospective examination of an athletic population. *Foot Ankle Int* 19: 653-660.
7. Hopkinson WJ, St Pierre P, Ryan JB, Wheeler JH (1990) Syndesmosis sprains of the ankle. *Foot Ankle* 10: 325-330.
8. Dattani R, Patnaik S, Kantak A, Srikanth B, Selvan TP (2008) Injuries to the tibiofibular syndesmosis. *J Bone Joint Surg Br* 90: 405-410.
9. Amendola A, Williams G, Foster D (2006) Evidence-based approach to treatment of acute traumatic syndesmosis (high ankle) sprains. *Sports Med Arthrosc* 14: 232-236.
10. Williams GN, Jones MH, Amendola A (2007) Syndesmotic ankle sprains in athletes. *Am J Sports Med* 35: 1197-1207.
11. Ogilvie-Harris DJ, Reed SC, Hedman TP (1994) Disruption of the ankle syndesmosis: biomechanical study of the ligamentous restraints. *Arthroscopy* 10: 558-560.
12. Xenos JS1, Hopkinson WJ, Mulligan ME, Olson EJ, Popovic NA (1995) The tibiofibular syndesmosis. Evaluation of the ligamentous structures, methods of fixation, and radiographic assessment. *J Bone Joint Surg Am* 77: 847-856.
13. KLEIGER B (1956) The mechanism of ankle injuries. *J Bone Joint Surg Am* 38-38A: 59-70.



14. Edwards GS Jr, DeLee JC (1984) Ankle diastasis without fracture. *Foot Ankle* 4: 305-312.
15. Turco VJ (1977) Injuries to the ankle and foot in athletics. *Orthop Clin North Am* 8: 669-682.
16. Fritschy D (1989) An unusual ankle injury in top skiers. *Am J Sports Med* 17: 282-285.
17. Magee DJ (1997) *Orthopedic Physical Assessment*. (3rd edn), WB Saunders. Philadelphia, USA
18. LAUGE-HANSEN N (1950) Fractures of the ankle. II. Combined experimental-surgical and experimental-roentgenologic investigations. *Arch Surg* 60: 957-985.
19. Pankovich AM (1976) Maisonneuve fracture of the fibula. *J Bone Joint Surg Am* 58: 337-342.
20. Boytim MJ, Fischer DA, Neumann L (1991) Syndesmotic ankle sprains. *Am J Sports Med* 19: 294-298.
21. den Daas A, van Zuuren WJ, Pelet S, van Noort A, van den Bekerom MP (2012) Flexible stabilization of the distal tibiofibular syndesmosis: clinical and biomechanical considerations: a review of the literature. *Strategies Trauma Limb Reconstr* 7: 123-129.
22. Zalavras C, Thordarson D (2007) Ankle syndesmotic injury. *J Am Acad Orthop Surg* 15: 330-339.
23. Rammelt S, Zwipp H, Grass R (2008) Injuries to the distal tibiofibular syndesmosis: an evidence-based approach to acute and chronic lesions. *Foot Ankle Clin* 13: 611-633, vii-viii.
24. van den Bekerom MP, Lamme B, Hogervorst M, Bolhuis HW (2007) Which ankle fractures require syndesmotic stabilization? *J Foot Ankle Surg* 46: 456-463.
25. van den Bekerom MP, Raven EE (2007) Current concepts review: operative techniques for stabilizing the distal tibiofibular syndesmosis. *Foot Ankle Int* 28: 1302-1308.
26. Kaukonen JP, Lamberg T, Korkala O, Pajarinen J (2005) Fixation of syndesmotic ruptures in 38 patients with a malleolar fracture: a randomized study comparing a metallic and a bioabsorbable screw. *J Orthop Trauma* 19: 392-395.
27. Thordarson DB, Samuelson M, Shepherd LE, Merkle PF, Lee J (2001) Bioabsorbable versus stainless steel screw fixation of the syndesmosis in pronation-lateral rotation ankle fractures: a prospective randomized trial. *Foot Ankle Int*: 335-338.
28. Yde J, Kristensen KD (1981) Inferior tibio-fibular diastasis treated by staple fixation. *J Trauma* 21: 483-485.
29. Marqueen T, Owen J, Nicandri G, Wayne J, Carr J (2005) Comparison of the syndesmotic staple to the transsyndesmotic screw: a biomechanical study. *Foot Ankle Int* 26: 224-230.
30. Relwani J, Lahoti O, Orakwe S (2002) Ilizarov ring fixator for a difficult case of ankle syndesmosis disruption. *J Foot Ankle Surg* 41: 335-337.
31. Thornes B, Walsh A, Hislop M, Murray P, O'Brien M (2003) Suture-endobutton fixation of ankle tibio-fibular diastasis: a cadaver study. *Foot Ankle Int* 24: 142-146.
32. Thornes B, Shannon F, Guiney AM, Hession P, Masterson E (2005) Suture-button syndesmosis fixation: accelerated rehabilitation and improved outcomes. *Clin Orthop Relat Res*: 207-212.
33. Farhan MJ, Smith TW (1985) Fixation of diastasis of the inferior tibiofibular joint using the syndesmosis hook. *Injury* 16: 309-311.
34. Engelbrecht E, Engelbrecht H, Huynh PL (1984) [Experiences with the syndesmosis hook in tibiofibular ligament injuries]. *Chirurg* 55: 749-755.
35. Grady JF, Moore CJ, O'Connor KJ, La Montagne D (1995) The use of a transsyndesmotic bolt in the treatment of tibiofibular diastasis: two case studies. *J Foot Ankle Surg* 34: 571-576.
36. Kabukcuoglu Y, Kucukkaya M, Eren T, Gorgec M, Kuzgun U (2000) The ANK device: a new approach in the treatment of the fractures of the lateral malleolus associated with the rupture of the syndesmosis. *Foot Ankle Int* 21: 753-758.
37. Kara AN, Esenyel CZ, Sener BT, Merih E (1999) A different approach to the treatment of the lateral malleolar fractures with syndesmosis injury: the ANK nail. *J Foot Ankle Surg* 38: 394-402.
38. Grass R, Rammelt S, Biewener A, Zwipp H (2003) Peroneus longus ligamentoplasty for chronic instability of the distal tibiofibular syndesmosis. *Foot Ankle Int* 24: 392-397.
39. Hovis WD, Kaiser BW, Watson JT, Bucholz RW (2002) Treatment of syndesmotic disruptions of the ankle with bioabsorbable screw fixation. *J Bone Joint Surg Am* 84-84A: 26-31.
40. Scranton PE Jr, McMaster JG, Kelly E (1976) Dynamic fibular function: a new concept. *Clin Orthop Relat Res*: 76-81.
41. Thordarson DB, Hedman TP, Gross D, Magre G (1997) Biomechanical evaluation of polylactide absorbable screws used for syndesmosis injury repair. *Foot Ankle Int* 18: 622-627.
42. Seitz WH Jr, Bachner EJ, Abram LJ, Postak P, Polando G, et al. (1991) Repair of the tibiofibular syndesmosis with a flexible implant. *J Orthop Trauma* 5: 78-82.
43. Burwell HN, Charnley AD (1965) The treatment of displaced fractures at the ankle by rigid internal fixation and early joint movement. *J Bone Joint Surg Br* 47: 634-660.
44. de Souza LJ, Gustilo RB, Meyer TJ (1985) Results of operative treatment of displaced external rotation-abduction fractures of the ankle. *J Bone Joint Surg Am* 67: 1066-1074.
45. Cottom JM, Hyer CF, Philbin TM, Berlet GC (2008) Treatment of syndesmotic disruptions with the Arthrex TightRope: a report of 25 cases. *Foot Ankle Int* 29: 773-780.
46. Thornes B, McCartan D (2006) Ankle Syndesmosis Injuries Treated with the TightRope Suture-Button Kit. *Techniques in Foot and Ankle Surgery*: 45-53
47. Coetzee JC, Ebeling P (2008) Treatment of Syndesmosis Distruptions with TightRope Fixation. *Techniques in Foot and Ankle Surgery*: 196-202
48. Soin SP, Knight TA, Dinah AF, Mears SC, Swierstra BA, et al. (2009) Suture-button versus screw fixation in a syndesmosis rupture model: a biomechanical comparison. *Foot Ankle Int* 30: 346-352.
49. Klitzman R, Zhao H, Zhang LQ, Strohmeyer G, Vora A (2010) Suture-button versus screw fixation of the syndesmosis: a biomechanical analysis. *Foot Ankle Int* 31: 69-75.
50. Degroot H, Al-Omari AA, El Ghazaly SA (2011) Outcomes of suture button repair of the distal tibiofibular syndesmosis. *Foot Ankle Int* 32: 250-256.
51. Teramoto A, Suzuki D, Kamiya T, Chikenji T, Watanabe K, et al. (2011) Comparison of different fixation methods of the suture-button implant for tibiofibular syndesmosis injuries. *Am J Sports Med* 39: 2226-2232.
52. Rigby RB, Cottom JM (2013) Does the Arthrex TightRope® provide maintenance of the distal tibiofibular syndesmosis? A 2-year follow-up of 64 TightRopes® in 37 patients. *J Foot Ankle Surg* 52: 563-567.
53. Schepers T (2012) *International Orthopaedics*. 36: 1199-1206.
54. Storey P, Gadd RJ, Blundell C, Davies MB (2012) Complications of suture button ankle syndesmosis stabilization with modifications of surgical technique. *Foot Ankle Int* 33: 717-721.
55. Espinosa N, Smerek JP, Myerson MS (2006) Acute and chronic syndesmosis injuries: pathomechanisms, diagnosis and management. *Foot Ankle Clin* 11: 639-657.
56. Karapinar H, Kalenderer O, Karapinar L, Altay T, Manisali M, et al. (2007) Effects of three- or four-cortex syndesmotic fixation in ankle fractures. *J Am Podiatr Med Assoc* 97: 457-459.