

## Autologous Particulate Dentin by Smart Dentin Grinder and Sorter

Itzhak Binderman\*


Department of Oral Biology, Tel Aviv University, Israel

\*Corresponding author: Itzhak Binderman, Department of Oral Biology, Faculty of Engineering, School of Dental Medicine, Tel Aviv University, Ramat Aviv 69978, Tel Aviv, Israel, Tel: +97-2505352344; E-mail: [binderman.itzhak@gmail.com](mailto:binderman.itzhak@gmail.com)

Rec date: Jul 14, 2015, Acc date: Jul 16, 2015, Pub date: Jul 20, 2015

Copyright: © 2015 Binderman I. This is an open-access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

### Clinical Images



**Figure 1:** (a) Smart Dentin Grinder and sorter (b) autologous particulate dentin after grinding and sorting; particle size of 300-1200 µm is collected and then grafted in extraction sites and bone defects.