

**Supplementary 1: Severity rating scale used at hospital.**

Rating	Description	Definition (Severity of Effect)
10	Dangerously high	Failure could injure the customer or an employee.
9	Extremely high	Failure would create noncompliance with federal regulations.
8	Very high	Failure renders the unit inoperable or unfit for use.
7	High	Failure causes a high degree of customer dissatisfaction.
6	Moderate	Failure results in a subsystem or partial malfunction of the product.
5	Low	Failure creates enough of a performance loss to cause the customer to complain.
4	Very Low	Failure can be overcome with modifications to the customer's process or product, but there is minor performance loss.
3	Minor	Failure would create a minor nuisance to the customer, but the customer can overcome it without performance loss.
2	Very Minor	Failure may not be readily apparent to the customer, but would have minor effects on the customer's process or product.
1	None	Failure would not be noticeable to the customer and would not affect the customer's process or product.

Source: FMEA Reference Guide. Quality Training Portal @ 2016 Resource Engineering, Inc.  
Available at: <http://www.qualitytrainingportal.com/resources/fmea/index.htm>

**Supplementary 2:** Occurrence Rating Scale used at hospital.

Rating	Description	Potential Failure Rate
10	Very High: Failure is almost inevitable.	More than one occurrence per day or a probability of more than three occurrences in 10 events.
9	High: Failures occur almost as often as not.	One occurrence every three to four days or a probability of three occurrences in 10 events.
8	High: Repeated failures.	One occurrence per week or a probability of 5 occurrences in 100 events.
7	High: Failures occur often.	One occurrence every month or one occurrence in 100 events.
6	Moderately High: Frequent failures.	One occurrence every three months or three occurrences in 1,000 events.
5	Moderate: Occasional failures.	One occurrence every six months to one year or five occurrences in 10,000 events.
4	Moderately Low: Infrequent failures.	One occurrence per year or six occurrences in 100,000 events.
3	Low: Relatively few failures.	One occurrence every one to three years or six occurrences in ten million events.
2	Low: Failures are few and far between.	One occurrence every three to five years or 2 occurrences in one billion events.
1	Remote: Failure is unlikely.	One occurrence in greater than five years or less than two occurrences in one billion events.


Source: FMEA Reference Guide. Quality Training Portal @ 2016 Resource Engineering, Inc.  
 Available at: <http://www.qualitytrainingportal.com/resources/fmea/index.htm>

**Supplementary 3: Detection Rating Scale used at hospital.**

Rating	Description	Definition
10	Absolute Uncertainty	The product is not inspected or the defect caused by failure is not detectable.
9	Very Remote	Product is sampled, inspected, and released based on Acceptable Quality Level (AQL) sampling plans.
8	Remote	Product is accepted based on no defectives in a sample.
7	Very Low	Product is 100% manually inspected in the process.
6	Low	Product is 100% manually inspected using go/no-go or other mistake-proofing gauges.
5	Moderate	Some Statistical Process Control (SPC) is used in process and product is final inspected off-line.
4	Moderately High	SPC is used and there is immediate reaction to out-of-control conditions.
3	High	An effective SPC program is in place with process capabilities ( $C_{pk}$ ) greater than 1.33.
2	Very High	All products are 100% automatically inspected.
1	Almost Certain	The defect is obvious or there is 100% automatic inspection with regular calibration and preventive maintenance of the inspection equipment.

Source: FMEA Reference Guide. Quality Training Portal @ 2016 Resource Engineering, Inc.  
 Available at: <http://www.qualitytrainingportal.com/resources/fmea/index.htm>

## Chi-Square Distribution Table



The shaded area is equal to  $\alpha$  for  $\chi^2 = \chi_{\alpha}^2$ .

<i>df</i>	$\chi_{.995}^2$	$\chi_{.990}^2$	$\chi_{.975}^2$	$\chi_{.950}^2$	$\chi_{.900}^2$	$\chi_{.100}^2$	$\chi_{.050}^2$	$\chi_{.025}^2$	$\chi_{.010}^2$	$\chi_{.005}^2$
1	0.000	0.000	0.001	0.004	0.016	2.706	3.841	5.024	6.635	7.879
2	0.010	0.020	0.051	0.103	0.211	4.605	5.991	7.378	9.210	10.597
3	0.072	0.115	0.216	0.352	0.584	6.251	7.815	9.348	11.345	12.838
4	0.207	0.297	0.484	0.711	1.064	7.779	9.488	11.143	13.277	14.860
5	0.412	0.554	0.831	1.145	1.610	9.236	11.070	12.833	15.086	16.750
6	0.676	0.872	1.237	1.635	2.204	10.645	12.592	14.449	16.812	18.548
7	0.989	1.239	1.690	2.167	2.833	12.017	14.067	16.013	18.475	20.278
8	1.344	1.646	2.180	2.733	3.490	13.362	15.507	17.535	20.090	21.955
9	1.735	2.088	2.700	3.325	4.168	14.684	16.919	19.023	21.666	23.589
10	2.156	2.558	3.247	3.940	4.865	15.987	18.307	20.483	23.209	25.188
11	2.603	3.053	3.816	4.575	5.578	17.275	19.675	21.920	24.725	26.757
12	3.074	3.571	4.404	5.226	6.304	18.549	21.026	23.337	26.217	28.300
13	3.565	4.107	5.009	5.892	7.042	19.812	22.362	24.736	27.688	29.819
14	4.075	4.660	5.629	6.571	7.790	21.064	23.685	26.119	29.141	31.319
15	4.601	5.229	6.262	7.261	8.547	22.307	24.996	27.488	30.578	32.801
16	5.142	5.812	6.908	7.962	9.312	23.542	26.296	28.845	32.000	34.267
17	5.697	6.408	7.564	8.672	10.085	24.769	27.587	30.191	33.409	35.718
18	6.265	7.015	8.231	9.390	10.865	25.989	28.869	31.526	34.805	37.156
19	6.844	7.633	8.907	10.117	11.651	27.204	30.144	32.852	36.191	38.582
20	7.434	8.260	9.591	10.851	12.443	28.412	31.410	34.170	37.566	39.997
21	8.034	8.897	10.283	11.591	13.240	29.615	32.671	35.479	38.932	41.401
22	8.643	9.542	10.982	12.338	14.041	30.813	33.924	36.781	40.289	42.796
23	9.260	10.196	11.689	13.091	14.848	32.007	35.172	38.076	41.638	44.181
24	9.886	10.856	12.401	13.848	15.659	33.196	36.415	39.364	42.980	45.559
25	10.520	11.524	13.120	14.611	16.473	34.382	37.652	40.646	44.314	46.928
26	11.160	12.198	13.844	15.379	17.292	35.563	38.885	41.923	45.642	48.290
27	11.808	12.879	14.573	16.151	18.114	36.741	40.113	43.195	46.963	49.645
28	12.461	13.565	15.308	16.928	18.939	37.916	41.337	44.461	48.278	50.993
29	13.121	14.256	16.047	17.708	19.768	39.087	42.557	45.722	49.588	52.336
30	13.787	14.953	16.791	18.493	20.599	40.256	43.773	46.979	50.892	53.672
40	20.707	22.164	24.433	26.509	29.051	51.805	55.758	59.342	63.691	66.766
50	27.991	29.707	32.357	34.764	37.689	63.167	67.505	71.420	76.154	79.490
60	35.534	37.485	40.482	43.188	46.459	74.397	79.082	83.298	88.379	91.952
70	43.275	45.442	48.758	51.739	55.329	85.527	90.531	95.023	100.425	104.215
80	51.172	53.540	57.153	60.391	64.278	96.578	101.879	106.629	112.329	116.321
90	59.196	61.754	65.647	69.126	73.291	107.565	113.145	118.136	124.116	128.299
100	67.328	70.065	74.222	77.929	82.358	118.498	124.342	129.561	135.807	140.169